

SSCOR S-SCORT III

Model 74000



Clearing The Airway Is Our #1 Priority

Operating Instructions & Maintenance Manual



SSCOR INC.

2125 N. Madera Rd. Unit C
Simi Valley, CA 93065 USA
Tel 1+818-504-4054

www.sscor.com

shop.sscor.com

Email: marketing@sscor.com

techsupport@sscor.com



Intertek



Clearing The Airway Is Our #1 Priority

Caution Notice

1. **S**SSCOR suction units are not designed or intended for use in extended procedures that require prolonged high vacuum/low airflow applications, as is the case in wound drainage or endoscopic use or in any other procedure that produces high vacuum levels within an occluded system for an extended period of time. Turn the suction unit off when it is not in use.
2. Federal law restricts this device to sale, distribution, and use by, or on the order of a physician, emergency medical technician, or other medical practitioner. For use by medical personnel trained in suctioning techniques and in the use of medical suction equipment.
3. This manual is restricted to the discussion of the use and maintenance of this device. It does not attempt to discuss professional techniques in suctioning procedures.
4. Operator should be thoroughly familiar with these operating instructions before this device is used.
5. Do not use in the presence of flammable agents or anesthetics.
6. The **S**-SCORT III produces a powerful vacuum. Do not use the **S**-SCORT III to suction neonates.
7. Before testing for vacuum over -300mmHg look for an expiration date on the canister (where applicable) and change the canister if the canister has passed the expiration date to minimize the possibility of implosion, which can occur when a canister is aged or damaged.
8. The suction pump must be reconnected to the charging source after each use and remain on charge until needed. Disconnect only for portable use. **Check the condition of the battery frequently**, see the battery test on page 6. **S**SSCOR's AC-DC charger only charges the battery, it will not power the suction device. The suction device will run on vehicle power when the unit is connected directly to the vehicle battery via the DC power cord.
9. Use of inappropriate chargers will result in suboptimal performance of the battery and possible destruction of the battery. Only use 80533.

**NOTE: DO NOT ATTEMPT TO CHANGE THE ELECTRICAL SYSTEM.
THIS UNIT OPERATES ON 12-14V DC ONLY**

S-SCORT III Model 74000

Battery Operated Portable Suction Pump

United States Patent No. 5134994

©2024 Duet, SSCOR SDC Catheter™ is a registered trademarks of SSCOR, Inc.



Clearing The Airway Is Our #1 Priority

Table of Contents

Caution-Notice	Page 2
General Description	Page 3
Important Features	Pages 4-5
Operating Instructions	Pages 6-7
Troubleshooting	Page 8
Warranty	Page 9
Maintenance	Pages 9-10
Disinfection	Page 10
General Specifications	Page 11
Replacing the Battery Pack	Page 12

General Description

The **S**-SCORT® III is a portable, 12V DC battery operated suction pump to be used by professional personnel trained in Emergency Care techniques for opening the airway by removing bodily fluids and particulate matter.

The Model 74000 is powered by sealed lead acid batteries, capable of driving the unit for 30 - 45 minutes. The batteries are recharged by direct connection to the vehicle, or by a AC/DC Dual Model Battery Charger (included with the unit). A sealed lead acid battery is a very stable and reliable battery. Many factors can affect the life of a battery:

- Leaving a unit switched on after there is no longer enough power to run the pump can cause a battery to deep discharge. This can reduce the life of, or destroy the battery.
- Failing to charge a battery for an extended period of time will also cause the battery to go into deep discharge.
- Low temperatures may reduce the available capacity.
- High temperatures may cause deformation of the battery case and damage the battery.

Sealed lead acid batteries can easily be maintained to permit proper operation of the equipment. To protect the battery, after each procedure turn the unit off, put the unit on charge and always store the pump at room temperature. The only way to assure the battery has functional capacity, even if it is indicated the battery is fully charged, is to perform the battery check suggested on page 6 of this operations manual. Replacement battery may be purchased at shop.sscor.com.


The **S**-SCORT III is designed to provide instant, effective suctioning, independent of external sources of power and can be pre-set to be activated immediately upon reaching the distressed patient, with no set up time required during the first few critical minutes of the code. Suction power is controllable for those instances when full power would be considered harmful to the patient. All controls are clearly labeled and easily accessible.

The **S**-SCORT III is equipped with a disposable collection canister which features a bacterial filter to screen airborne particulates and a mechanical shut-off valve to prevent fluid overflow.



Clearing The Airway Is Our #1 Priority

Important Features





Clearing The Airway Is Our #1 Priority

Important Features

<u>Key #</u>	<u>Component</u>	<u>Function</u>
1	Vacuum Line	Connects regulator to canister
2	Vacuum Port	Connects canister to vacuum line
3	Patient Port	Connects canister to patient tube
4	Patient Connecting Tubing	9/32" ID, SSCOR part #43200
5	Disposable Canister	1200cc/ml, SSCOR part #48041
6	SSCOR Suction Catheter	SSCOR SDC Catheter, SSCOR part #200-00002 Or SSCOR HI-D Catheter, SSCOR part #44241
7	Charging Indicator Light	Signifies electrical connection from charger to the unit
8	On/Off Switch	Turns unit on or off
9	Charging Receptacle	To connect charger to battery
10	Carrying Handle	
11	Regulator Control	Adjusts negative pressure
12	Battery Connection	Quick disconnect
13	Pump	Provides vacuum source
14	Battery	Supplies electrical power to pump
15	Battery Charger	Charges battery after each use
16	12V DC Electrical Plug	To connect charger to battery
17	DC Power Cord	Connects power cord to vehicle



Clearing The Airway Is Our #1 Priority

Operating Instructions

For optimal performance the suction device is to remain plugged in whenever the device is not in use. If for any reason poor battery quality is suspected, a battery test should be conducted.

Battery Test

Run the following test whenever poor battery quality is suspected to ensure proper performance of the device.

1. Confirm the power cord is supplying power to the device. Check the power indicator light on the control panel.
2. Remove the power cord and run the unit from its internal DC battery.
3. Check for vacuum by occluding the patient tube and set the vacuum regulator to the maximum vacuum setting. Un-occlude the patient tubing.
4. Allow the unit to run for 15 minutes on DC power. If the unit stops or slows during the 15 minutes, it is possible the battery capacity has been depleted. It is time to replace the battery.
5. If the unit is still running at full power after 15 minutes, adjust the regulator to the desired setting, turn the device off and put it back on charge.
6. Reconnect the unit to the charging source as soon as possible after each use.

SSCOR recommends replacing the battery after 3 years. Replacement batteries may be purchased at shop.sscor.com.

Battery Charging Suggestions

It is important to keep the battery connected to the charging source at all times when the unit is not in use. In order for the **S-SCORT® III** battery to operate at its maximum efficiency, it must be charged from a power source that will bring the voltage level of the battery up to the 13V DC operating range. This **S-SCORT III** model is equipped to charge the battery two ways, with a AC/DC battery charger (15) or a DC power cord for direct connection to the vehicle (17).

An active vehicle, running calls around the clock, will do well to charge the **S-SCORT III** from the vehicle's DC charging system. The preferred method of charging the **S-SCORT III** is to wire the DC power cord to the vehicle DC power system on a properly fused line in front of the master switch. This charging method is designed to keep the battery charging at all times. If the suction unit is operated while it is hooked up to the vehicle it will utilize the vehicle power and save its own battery for emergency use. All **SSCOR, Inc.** suction units have a diode to prevent drawdown from the pump to the vehicle electrical system and a fuse to protect the pump from vehicle electrical surges. If your suction unit is wired to the vehicle battery via an automatic load switch power supply, be sure to use a filter in order to eliminate any voltage spikes.

A vehicle running infrequent calls does not charge itself often enough to keep batteries in the 13V DC operating range and should rely on the **S-SCORT III** AC/DC battery charger when the vehicle is connected to the shore line. In this instance the DC power cord should be used as a backup source of power to run the **S-SCORT III** from the vehicle during patient transport. To utilize the 115V AC battery charger, the suction pump must be removed from the vehicle and charged from an AC outlet or the charger must be plugged into a "shore-line" outlet in the vehicle. Whichever charging method is chosen, remember to keep the unit connected to the charging source whenever the unit is not in use.

For replacement parts and accessories, please visit shop.sscor.com



Clearing The Airway Is Our #1 Priority

Operating Instructions

NOTICE: Component key numbers are indicated within ()s. Reference pages 4 and 5.

Electrical operation: Constant charging is required.

CHARGING THE S-SCORT III FROM A VEHICLE

Connect the DC power cord to the cigarette lighter receptacle in the vehicle or hard wire the DC power cord (17) to the hot DC electrical system of the vehicle on a properly fused line in front of the master switch.

Connect the electrical line cord to the S-SCORT® III by securely attaching the charging plug (16) into the receptacle (9). The S-SCORT III battery is charged by the vehicle electrical system and the pump is powered by the vehicle current and the S-SCORT III battery running in parallel.

DC CHARGING FROM AC/DC DUAL MODE BATTERY CHARGER

If 115V AC charging is required, connect the S-SCORT III to a SSCOR AC battery charger (15) SSCOR Part 80533 (included with unit). AC charging should bring the battery to a dependable working charge in 4 to 6 hours. The green LED on the charger will light when the battery is fully charged.

The charging indicator light (7) on the front panel indicates the S-SCORT III is receiving electrical input from the AC charger or vehicle. It does not reflect the condition of the battery. Check the charging Indicator light to be sure you have a good electrical connection.

OPERATION OF UNIT

Negative pressure on the Model 74000 is controlled by a two-position regulator (11). When fully depressed (pushed toward the pump), the negative pressure exceeds -525mmHg. To reduce negative pressure, pull the regulator straight out to the stop. In this position, the negative pressure will be -120mmHg ($\pm 15\%$). Attach the patient connecting tube (4) to the patient port (3) on the canister (5). Use 9/32" I.D. tubing. Turn the On/Off switch (8) to the On position and occlude the end of the patient connecting tube. Keep it occluded for 10 seconds. Release the occlusion and observe evidence of negative pressure. If the pump is running and no negative pressure is observed, check to be sure the lid on the disposable canister is tight and vacuum connections are secure. Dispose of the canister after use according to local / regional / national requirements for the disposal of hazardous waste materials.

Reconnect the S-SCORT III to the vehicle or to the battery charger (15) as soon as possible following the code by securely attaching charging plug (16) into the receptacle on the unit (9).

CAUTION - Always turn the pump switch "Off" as soon as possible after the procedure. Rechargeable batteries will short out if they continue to discharge after they have reached the point where they do not have sufficient power to drive the pump. Be sure the switch is in the "Off" position until needed.



Clearing The Airway Is Our #1 Priority

Troubleshooting

PROBLEM	POSSIBLE CAUSE	CORRECTIVE ACTION
Unit does not function when switch is in the "ON" position	· Molex connections disconnected · Battery discharged · Battery damaged · Charger damaged	· Reconnect Molex connections · Recharge battery · Replace battery · Check output of charger
Unit does not suction when pump is running	· Vacuum line loose · Canister damaged · Canister lid is not tight · Thumb vent on suction tip is not occluded	· Check connections · Replace canister · Re-secure canister lid · Occlude thumb vent with thumb
Pump is sluggish	· Residual materials have collection the pump head · Battery is unable to retain a charge · Loose connections	· Replace pump · Replace battery · Check connections
System shuts down while suctioning heavy particulate matter	· Vacuum line clogged at canister lid · Float valve has closed	· Remove connector or canister lid and loosen obstruction · Loosen float valve, empty contents, or replace canister
Regulator stem is difficult to pull	· Stem requires lubrication	· Lubricate stem with pneumatics lubricant



Clearing The Airway Is Our #1 Priority

Warranty

SSCOR warrants that each new product is free from defects in material and workmanship under normal use and service for a period of five years from date of original purchase from SSCOR, Inc. If returned to SSCOR, we will arrange for repairs or replacement within the terms of the warranty. The product should be decontaminated and returned properly packaged and postage prepaid. Loss or damage in transit to the factory shall be at the purchaser's risk. Loss or damage of return shipment from SSCOR shall be at the purchaser's risk. Please go to the support page at [www. SSCOR.com](http://www.SSCOR.com) to request a return authorization number.

Conditions:

This warranty covers none of the following:

1. The warranty shall not apply to any SSCOR device in which a component has been altered; determined to be using a non SSCOR component; or altered in any way so as, in SSCOR's judgment, to affect its safety or efficacy.
2. SSCOR product which has been subject to misuse, negligence, or accident.
3. The serial number has been altered, effaced or removed.
4. SSCOR product which has been setup otherwise than in accordance with the instructions furnished by SSCOR.
5. Maintenance and repair due to normal wear and tear.
6. Installation of product in a way not shown in instruction for use furnished by SSCOR.
7. Faults in the system which SSCOR product is installed.
8. Batteries, Disposables items (canisters, patient tubing, and catheters), replacement parts, cosmetic irregularities that do not degrade the safety and efficacy of the SSCOR product.

This warranty is in lieu of all other warranties expressed or implied and of all other obligations or liabilities on SSCOR's part, and SSCOR neither assumes, nor authorizes any representative or other persons to assume for it, any other liability in connection with the sale of SSCOR products.

This warranty gives you specific legal rights and you may also have other rights that vary from jurisdiction to jurisdiction. For countries where minimum warranty terms are determined by statute, the warranty term is the longer of the statutory period or the term listed above

Maintenance

Preventive Care: Observe the following maintenance routine to ensure readiness at any time:

1. When the SSCOR aspirator is not in use, keep batteries on continuous charge.
2. Test the SSCOR aspirator at regular intervals; See page 6.
3. Make sure the SSCOR aspirator is always clean and ready for use.
4. If the procedure produced an excessive quantity of fluids, check the vacuum line (1) for evidence of moisture. If the vacuum line between the pump and canister is moist, it is possible that fluids have reached the vacuum pump. See Disinfection Instructions (page 10).
5. If the vacuum pump appears defective, return the unit to the factory for repair. Do not attempt to repair the vacuum pump.
6. For technical assistance, call +1 (818) 504-4054. For replacement parts and accessories, please visit shop.sscor.com.

Maintenance

As soon as possible after use, the single use disposable canister, patient tubing and catheter should be discarded according to local / regional / national requirements for the disposal of hazardous waste materials. Clean the exterior of the SSCOR suction unit using a mild detergent and clear water by dampening a clean lint free cloth. Rinse using clear water and another damp clean lint free cloth to remove any detergent residue.

NOTE: The hydrophobic filter in the canister helps to ensure that no moisture or particulate matter reaches the inside of the device. When fluids fill the canister, the positive (mechanical float) shutoff valve closes immediately, shutting the vacuum port off so as to prevent fluid from contacting the pump. The filter has been tested by the manufacturer (Bemis) to screen out aerosolized microorganisms and particulate matter at a bacterial efficiency rating of 99.99% DOP. The canister also has sidewall gradation marks starting at 100 ml/cc and at every 50 ml/cc up to 1200 ml/cc indicating the fill level of the canister.

In the unlikely event that fluids may have reached the vacuum pump, read the disinfection section. Your engineering department will have to open the unit to check the condition of the pump. Do not reuse any single use disposable parts; do not submerge the device into any liquid, this will void the warranty and cause the device to malfunction.

Disinfection

Use personal protective equipment such as gloves, a smock, and face and eye protection when handling units that are suspected to be contaminated.

Part	Cleaning and Disinfecting
Collection Canister	Disposable item, re-use not permitted. Use new canister for each patient.
Patient Tubing	Disposable item, re-use not permitted. Use new patient tubing for each patient.
SSCOR Suction Catheter	Disposable item, re-use not permitted. Use new a new SSCOR Suction Catheter for each patient.
Vacuum Pump	Wipe with damp cloth or disinfectant wipe. Sterilization not permitted. Vacuum pump should be replaced if contaminated
Chassis	Wipe with damp cloth or disinfectant wipe. Sterilization not permitted.

Caution: Disconnect the unit from any power source prior to cleaning the unit.

Disinfect the unit using a mild surface disinfectant, such as a 10:1 mixture of water and bleach. The unit is designed to suction contaminated fluids, which should be removed from the system immediately after use. In the unlikely event that fluids may have reached the vacuum pump, your engineering department will have to open the unit to check the condition of the pump. When cleaning the interior of the chassis, disconnect the battery from the wiring harness. The only foreseeable way fluids may reach the vacuum pump is the filter in the canister has been compromised or bypassed.

For technical assistance, call +1 (818) 504-4054. For replacement parts and accessories, please visit shop.sscor.com.



Clearing The Airway Is Our #1 Priority

General Specifications

CHARACTERISTICS	SPECIFICATIONS
Size	11"L x 8"H x 5.25"W
Weight	7 pounds
Vacuum Pump	12V DC., 3.0 A Clinical Airflow $\geq$ 30LPM Exceeds 525mmHg
Regulator	Controls negative pressure
Power Source: Battery (DC Power) DC Charging from Vehicle	Rechargeable Sealed Lead Acid. SSCOR part # 80638. Please read pages 6 and 7 for battery care suggestions. At full capacity a fully charged battery will power the suction device for 30-45 minutes. Vehicle electrical system powers the pump and charges the battery until disconnected and then the unit switches automatically to its self contained battery.
AC Charger	U.L. Listed AC/DC Dual Mode Battery Charger
Switch	On/Off Rocker
Collection Canister	1200cc/ml SSCOR part #48041
Patient tubing	Vinyl tubing 9/32"ID x 72"L SSCOR part #43200
SSCOR Suction Catheter	SSCOR SDC Catheter, SSCOR part #200-00002 Or SSCOR HI-D Catheter, SSCOR part #44241



Clearing The Airway Is Our #1 Priority

Replacing the Battery. (Available for purchase on shop.sscor.com)

Refer to parts diagram pages 4 and 5.

Instructions to replace the battery

- Unzip the top flap and pull it back exposing the foam chassis.
- Remove the top portion of the foam chassis exposing the components.
- Lift the battery out of the chassis and disconnect the battery's Molex fitting from the wiring harness.
- Dispose of the old battery according to local / regional / national requirements for the disposal of sealed lead acid batteries.

Instructions for disassembly to wash the fabric case

- Remove the four screws around the outside switch plate and remove the plate.
- Unzip the top flap and pull it back exposing the foam chassis.
- Remove the top portion of the foam chassis exposing the components.
- Remove the screw stop that attaches to the white regulator stem and remove the regulator stem.
- Use a flat head screwdriver to remove the "E" clip from the outside of the regulator.
- Pull the vacuum tubing back to the pump head, under the flap.
- All of the components should lift out of the chassis and remove the foam chassis from the fabric case.

To reassemble the device

- Put the foam chassis back into the fabric case.
- Reattach the switch plate with the four screws.
- Put the pump and battery back into the foam chassis.
- Replace the white regulator stem into the regulator. Rotate the stem and place the set screw through the regulator slot & into the set screw hole on the white regulator stem. Do not screw all of the way down (approx. 1/16" from regulator stem). Check the regulator stem moves freely.
- Pull the stem all of the way out & place the "E" clip onto the end of the brass regulator (on the outside of the case) using channel locks to snap the "E" clip into place.
- Reattach the vacuum tubing to the vacuum port of the canister. Close the zipper on the top of the case. Turn the unit on and verify there is a vacuum.